



Testing Laboratory
**Accreditation
 Certificate**

Accreditation No. RTL02510



NIHON DEMPA KOGYO CO., LTD.
Quality Assurance Laboratory

1275-2, Kamihirose, Sayama-shi, Saitama, 350-1321 Japan

meets the following criteria. On the basis of this, Japan Accreditation Board (JAB) grants accreditation to the said testing laboratory.

- Applicable accreditation criteria : ISO/IEC 17025:2017 (JIS Q 17025:2018)
- Scope of accreditation : **Electrical Testing**
(As described in the appendix)
- Premises covered by accreditation : As described in the appendix.
- Expiry date of accreditation : February 29, 2028

	Revised	April 22, 2026
	Renewed	March 1, 2024
	Initial accreditation	February 4, 2008

Y. Miki, President

Japan Accreditation Board



Accreditation No.

RTL02510

Accreditation Certificate

Appendix

(Page 1/2)

Type of Laboratory	Testing
Name of Laboratory	NIHON DEMPA KOGYO CO., LTD. Quality Assurance Laboratory
Address	1275-2, Kamihirose, Sayama-shi, Saitama, 350-1321 Japan

1) Premises on which testing activities are performed

Name of Premises	NIHON DEMPA KOGYO CO., LTD. Quality Assurance Laboratory
Address	1275-2, Kamihirose, Sayama-shi, Saitama, 350-1321 Japan
Testing service at permanent facilities or on site testing service	<input checked="" type="checkbox"/> Testing service at permanent facilities <input type="checkbox"/> On site testing service

Scope of Accreditation

FIELD	M21 Electrical Testing
-------	------------------------

CODE OF CLASSIFICATION, NAME	TEST METHOD STANDARD
M21.5 Environmental testing M21.5.1 Cold tests	JIS C 60068-2-1:1995, IEC 60068-2-1:1990 + A1:1993 + A2:1994 Test Aa(except 6,8,10) Test temperature: -40 °C JIS C 60068-2-1:2010, IEC 60068-2-1:2007 Test Ab(except 6,8,6.10,6,13) Test temperature: -40 °C
M21.5 Environmental testing M21.5.2 Dry heat tests	JIS C 60068-2-2:1995, IEC 60068-2-2:1974 + A1:1993 Test Ba(except 6,8,10) Test temperature: 85 °C, 100 °C, 125 °C, 155 °C, 175 °C
M21.5 Environmental testing M21.5.3 Change of temperature	JIS C 0025:1988 Test Na(except 2.5,2.8 and except 2.6.2 the exposure time 10 min) IEC 60068-2-14:1984 Test Na(except 1.5,1.8 and except 1.6.2 the exposure time 10 min) Test temperature : -40 °C, 85 °C, 100 °C, 125 °C, 155 °C
M21.5 Environmental testing M21.5.5 Damp heat tests, steady state	JIS C 60068-2-78:2004(except 6,8,10), IEC 60068-2-78:2001(except 6,8,10) Test combination: 40 °C and 93 %RH



Accreditation No.

RTL02510

Accreditation Certificate

Appendix

(Page 2/2)

Type of Laboratory	Testing
Name of Laboratory	NIHON DEMPA KOGYO CO., LTD. Quality Assurance Laboratory
Address	1275-2, Kamihirose, Sayama-shi, Saitama, 350-1321 Japan

(Notes on Accreditation Certificate)

The laboratory is only accredited for laboratory activities outlined within the methods listed above. Reference to any other activity within these standards, such as risk management or risk assessment, does not fall within the laboratory's accredited capabilities.

When version information of standards or methods are not identified in the scope, laboratories shall adapt to use the current version of such standards within six months at latest from the issued date of current version.

Notes for EMC test laboratory for FCC

Accreditation does not imply acceptance to the FCC equipment authorization program. Please see the FCC website (<https://apps.fcc.gov/oetcf/eas/>) for a listing of FCC approved laboratories.

Japan Accreditation Board